
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000515
13:03 21/05/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:03
PRESSURIZING 3:08
STERILIZING 4:16
136.0 C 229kPa 4:16
135.9 C 229kPa 4:46
136.3 C 232kPa 5:16
136.4 C 232kPa 5:46
136.4 C 233kPa 6:16
136.1 C 231kPa 6:46
136.3 C 233kPa 7:16
136.3 C 233kPa 7:46
136.3 C 233kPa 7:46
VENTING 7:46
AIR DRYING 8:42
CYCLE COMPLETE 23:42

Digital Signature #
79E85568E89EB9C5
