
Midmark M11 - v1.0.0
Total Cycles: 1736

Name: RITTER 1
Sterilizer ID: V2178198
03/23/2022 01:16 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 210.1 F 0.0
2:00 203.6 F 3.3
4:00 207.8 F 8.9
6:00 223.7 F 7.2
8:00 234.5 F 10.7
10:00 234.3 F 8.0
12:00 241.4 F 10.4
14:00 247.6 F 14.0
16:00 260.8 F 21.3
18:00 272.3 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.4 F 28.7
2:00 272.1 F 29.0
3:00 272.0 F 28.4
4:00 272.5 F 29.2

Min 271.4 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 216.0 F 0.0
5:00 233.2 F 0.0
10:00 229.0 F 0.0
15:00 228.2 F 0.0
20:00 226.8 F 0.0
25:00 230.0 F 0.0
30:00 229.8 F 0.0
Drying Complete

FILLING: 2:21
HEATING: 18:25
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 00:57:50
CYCLE COMPLETE

03/23/2022 02:13 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____