

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4278

Name: RITTER 1  
Sterilizer ID: V2178198  
02/07/2024 11:08 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 142.4 F 0.0  
2:00 139.7 F 1.3  
4:00 144.2 F 3.7  
6:00 165.2 F 6.7  
8:00 185.9 F 10.5  
10:00 221.4 F 5.4  
12:00 227.5 F 7.0  
14:00 232.8 F 11.0  
16:00 231.4 F 6.9  
18:00 241.6 F 11.6  
20:00 237.6 F 9.1  
22:00 249.6 F 14.8  
24:00 260.0 F 20.6  
26:00 269.7 F 26.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.0  
1:00 272.7 F 28.6  
2:00 271.1 F 27.9  
3:00 272.7 F 29.0  
4:00 271.6 F 28.5

Min 271.0 F 27.9  
Max 273.1 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 183.7 F 0.0  
5:00 232.5 F 0.0  
10:00 187.2 F 0.0  
15:00 183.4 F 0.0  
20:00 181.5 F 0.0  
25:00 180.7 F 0.0  
30:00 179.0 F 0.0  
Drying Complete

FILLING: 0:55  
HEATING: 27:03  
STERILIZING: 4:00  
VENTING: 3:43  
DRYING: 30:00  
TOTAL CYCLE: 01:05:40  
CYCLE COMPLETE

02/07/2024 12:13 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_