
Midmark M11 - v1.0.5
Total Cycles: 981

Name: RITTER2
Sterilizer ID: V2313228
04/13/2022 09:10 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	136.0 F	0.0
2:00	136.0 F	0.0
4:00	147.9 F	0.2
6:00	184.5 F	0.6
8:00	204.7 F	4.2
10:00	218.6 F	8.7
12:00	230.6 F	11.9
14:00	235.1 F	10.2
16:00	236.2 F	8.7
18:00	241.8 F	11.1
20:00	247.3 F	14.2
22:00	259.8 F	21.2
24:00	271.0 F	28.5

STERILIZING

mm:ss	Degrees	PSI
0:00	272.1 F	28.9
1:00	272.6 F	29.6
2:00	271.7 F	28.7
3:00	272.7 F	29.6
4:00	271.8 F	28.8

Min 271.5 F 28.7
Max 272.7 F 29.7

VENTING CHAMBER

mm:ss	Drying Start	Degrees	PSI
0:00		207.7 F	0.0
5:00		200.9 F	0.0
10:00		205.8 F	0.0
15:00		218.5 F	0.0
20:00		225.4 F	0.0
25:00		229.4 F	0.0
30:00		229.8 F	0.0

Drying Complete

FILLING: 0:51
HEATING: 24:35
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:02:34
CYCLE COMPLETE

04/13/2022 10:12 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____