

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1856

Name: RITTER 1  
Sterilizer ID: V2178198  
04/22/2022 09:44 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 71.9 F 0.0  
2:00 78.0 F 0.0  
4:00 115.2 F 0.1  
6:00 148.6 F 1.6  
8:00 168.7 F 4.8  
10:00 189.4 F 9.2  
12:00 207.5 F 11.4  
14:00 219.9 F 5.7  
16:00 231.9 F 9.7  
18:00 236.5 F 9.8  
20:00 241.2 F 11.0  
22:00 236.5 F 8.5  
24:00 252.0 F 15.9  
26:00 264.6 F 23.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 27.8  
1:00 272.3 F 28.4  
2:00 272.8 F 28.5  
3:00 272.4 F 28.1  
4:00 271.9 F 27.8

Min 271.6 F 27.6  
Max 272.8 F 28.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 220.3 F 0.0  
5:00 232.2 F 0.0  
10:00 223.9 F 0.0  
15:00 216.7 F 0.0  
20:00 210.3 F 0.0  
Dry Cycle Aborted

FILLING: 1:01  
HEATING: 27:40  
STERILIZING: 4:00  
VENTING: 2:34  
DRYING: 20:34  
TOTAL CYCLE: 00:55:48

04/22/2022 10:39 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_