
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002003
14:30 12/07/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:32
PRESSURIZING 3:33
STERILIZING 4:38
136.4 C 221kPa 4:38
136.3 C 224kPa 5:08
136.4 C 222kPa 5:38
136.2 C 223kPa 6:08
136.4 C 224kPa 6:38
136.4 C 225kPa 7:08
136.4 C 223kPa 7:38
136.3 C 225kPa 8:08
136.3 C 225kPa 8:08
VENTING 8:08
AIR DRYING 9:04
CYCLE COMPLETE 9:05

Digital Signature #
EF499654CAE99A49
