
Midmark M11 - v1.0.5
Total Cycles: 4229

Name: RITTER2
Sterilizer ID: V2313228
12/19/2024 09:49 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.9 F 0.0
2:00 85.5 F 0.0
4:00 127.2 F 0.0
6:00 110.1 F 0.1
8:00 146.7 F 0.9
10:00 175.3 F 3.8
12:00 200.2 F 7.4
14:00 216.5 F 11.3
16:00 218.0 F 4.7
18:00 232.0 F 8.9
20:00 241.0 F 13.3
22:00 238.6 F 10.4
24:00 237.1 F 8.5
26:00 249.3 F 15.0
28:00 260.8 F 21.5
30:00 271.5 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.6 F 29.0
2:00 272.0 F 29.0
3:00 272.3 F 28.8
4:00 272.2 F 29.2

Min 271.5 F 28.4
Max 272.9 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.9 F 0.0
5:00 220.1 F 0.0
10:00 232.3 F 0.0
15:00 227.9 F 0.0
20:00 234.8 F 0.0
25:00 230.5 F 0.0
30:00 217.6 F 0.0
Drying Complete

FILLING: 1:00
HEATING: 30:37
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:09:02
CYCLE COMPLETE

12/19/2024 10:57 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____