

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
1.3uS / 0.8ppm  
CYCLE NUMBER 000607  
15:16 26/08/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:50  
PRESSURIZING 3:53  
STERILIZING 5:11  
136.0 C 228kPa 5:11  
136.4 C 233kPa 5:41  
136.4 C 232kPa 6:11  
136.2 C 229kPa 6:41  
136.4 C 233kPa 7:11  
136.4 C 231kPa 7:41  
136.2 C 230kPa 8:11  
136.3 C 232kPa 8:41  
136.3 C 232kPa 8:41  
VENTING 8:41  
AIR DRYING 9:38  
CYCLE COMPLETE 24:38  
-----

Digital Signature #  
349D749DDE650EFF  
-----