

-----  
Midmark M11 - v1.0.5  
Total Cycles: 24

Name: RITTER2  
Sterilizer ID: V2313228  
04/27/2021 09:39 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 146.6 F 0.0  
2:00 145.4 F 2.2  
4:00 161.7 F 5.6  
6:00 186.6 F 10.0  
8:00 209.8 F 5.6  
10:00 221.9 F 9.0  
12:00 232.7 F 10.5  
14:00 238.9 F 12.0  
16:00 239.7 F 10.9  
18:00 252.7 F 17.5  
20:00 264.3 F 24.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.3  
1:00 272.6 F 29.8  
2:00 271.5 F 29.4  
3:00 272.6 F 29.9  
4:00 271.6 F 29.2  
  
Min 271.4 F 29.2  
Max 272.7 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 181.8 F 0.0  
5:00 212.8 F 0.0  
10:00 207.0 F 0.0  
15:00 208.8 F 0.0  
20:00 205.9 F 0.0  
25:00 203.7 F 0.0  
30:00 201.6 F 0.0  
Drying Complete

FILLING: 1:14  
HEATING: 21:58  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 01:00:21  
CYCLE COMPLETE

04/27/2021 10:39 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_