
Midmark M11 - v1.0.0
Total Cycles: 4784

Name: RITTER 1
Sterilizer ID: V2178198
08/02/2024 12:10 PM

BEGIN UNWRAPPED CYCLE
Temp: 270 Degrees F
Time: 3 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.9 F 0.0
2:00 91.7 F 0.0

C103: HEATUP MODE
STOP PRESSED
CHAMBER TEMP 112F
CHAMBER PRESS 0.0PSI
** ITEMS NOT STERILE **

08/02/2024 12:13 PM