
Midmark M11 - v1.0.5
Total Cycles: 233

Name: RITTER2
Sterilizer ID: V2313228
07/21/2021 01:21 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.2 F 0.0
2:00 190.4 F 4.0
4:00 199.2 F 9.6
6:00 219.3 F 7.4
8:00 231.2 F 11.1
10:00 235.1 F 9.3
12:00 239.3 F 9.9
14:00 248.7 F 15.1
16:00 260.4 F 21.7
18:00 270.9 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.5
1:00 271.8 F 29.7
2:00 271.9 F 30.0
3:00 271.9 F 30.0
4:00 271.7 F 29.8

Min 271.5 F 29.4
Max 272.4 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.4 F 0.0
5:00 231.0 F 0.0
10:00 233.1 F 0.0
15:00 227.0 F 0.0
20:00 225.3 F 0.0
25:00 223.6 F 0.0
Dry Cycle Aborted

FILLING: 2:15
HEATING: 19:23
STERILIZING: 4:00
VENTING: 3:20
DRYING: 27:09
TOTAL CYCLE: 00:56:07

07/21/2021 02:17 PM
ACCEPTED: X
REJECTED:
Operator: U01