
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 000601
9:02 26/08/2021
WRAPPED
135 C for 10 min.
CYCLE START 0:00
CONDITIONING 2:15
PRESSURIZING 4:30
STERILIZING 5:30
136.0 C 229kPa 5:30
136.4 C 232kPa 6:00
136.4 C 233kPa 6:30
136.2 C 232kPa 7:00
136.1 C 232kPa 7:30
136.3 C 232kPa 8:00
136.2 C 230kPa 8:30
136.4 C 233kPa 9:00
136.3 C 232kPa 9:30
136.2 C 231kPa 10:00
136.4 C 233kPa 10:30
136.2 C 231kPa 11:00
136.1 C 231kPa 11:30
136.4 C 233kPa 12:00
136.4 C 233kPa 12:30
136.4 C 233kPa 13:00
136.3 C 232kPa 13:30
136.4 C 233kPa 14:00
136.3 C 230kPa 14:30
136.4 C 233kPa 15:00
136.4 C 231kPa 15:30
136.4 C 231kPa 15:30
VENTING 15:30
AIR DRYING 16:20
CYCLE COMPLETE 31:20

Digital Signature #
B01B31B27D9F54A3
