
Midmark M11 - v1.0.5
Total Cycles: 1318

Name: RITTER2
Sterilizer ID: V2313228
07/19/2022 10:26 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 155.5 F 0.0
2:00 161.6 F 0.1
4:00 196.9 F 2.7
6:00 217.5 F 8.0
8:00 233.0 F 13.4
10:00 232.2 F 9.5
12:00 239.2 F 9.6
14:00 242.3 F 13.5
16:00 243.2 F 12.8
18:00 256.0 F 19.5
20:00 267.4 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.5
1:00 272.7 F 30.1
2:00 271.8 F 29.5
3:00 272.9 F 30.2
4:00 272.1 F 29.5

Min 271.7 F 29.4
Max 273.0 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.5 F 0.0
5:00 198.0 F 0.0
10:00 214.7 F 0.0
15:00 224.8 F 0.0
20:00 226.1 F 0.0
25:00 225.1 F 0.0
30:00 225.9 F 0.0
Drying Complete

FILLING: 1:30
HEATING: 21:25
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:00:07
CYCLE COMPLETE

07/19/2022 11:26 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____