

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1001

Name: RITTER 1  
Sterilizer ID: V2178198  
07/13/2021 09:02 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 172.6 F 0.0  
2:00 174.8 F 2.3  
4:00 180.2 F 5.8  
6:00 209.3 F 10.8  
8:00 220.5 F 5.1  
10:00 232.4 F 10.4  
12:00 236.1 F 7.6  
14:00 243.5 F 11.9  
16:00 245.5 F 12.8  
18:00 258.1 F 19.5  
20:00 269.2 F 26.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.1  
1:00 272.6 F 28.8  
2:00 271.5 F 28.0  
3:00 272.7 F 28.9  
4:00 272.0 F 28.4  
  
Min 271.4 F 28.0  
Max 272.7 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.2 F 0.0  
5:00 213.1 F 0.0  
10:00 206.6 F 0.0  
15:00 207.5 F 0.0  
20:00 223.5 F 0.0  
25:00 229.6 F 0.0  
30:00 230.4 F 0.0  
Drying Complete

FILLING: 1:17  
HEATING: 21:04  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:59:31  
CYCLE COMPLETE

07/13/2021 10:06 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01