
Midmark M11 - v1.0.0
Total Cycles: 1502

Name: RITTER 1
Sterilizer ID: V2178198
01/10/2022 09:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 148.5 F 0.0
2:00 152.6 F 2.1
4:00 161.1 F 5.7
6:00 182.7 F 10.3
8:00 209.7 F 3.4
10:00 223.3 F 8.8
12:00 234.9 F 10.2
14:00 239.1 F 10.9
16:00 235.9 F 8.7
18:00 250.3 F 15.4
20:00 261.8 F 22.0
22:00 272.4 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 271.8 F 28.4
2:00 272.5 F 28.8
3:00 272.4 F 29.1
4:00 271.5 F 28.4

Min 271.4 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.4 F 0.0
5:00 209.5 F 0.0
10:00 226.3 F 0.0
15:00 226.0 F 0.0
20:00 229.2 F 0.0
25:00 230.2 F 0.0
30:00 229.1 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 22:22
STERILIZING: 4:00
VENTING: 2:56
DRYING: 30:00
TOTAL CYCLE: 01:00:11
CYCLE COMPLETE

01/10/2022 10:59 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____