

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 001090  
13:31 12/01/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:35  
PRESSURIZING 3:35  
STERILIZING 4:41  
136.4 C 222kPa 4:41  
136.2 C 218kPa 5:11  
136.3 C 224kPa 5:41  
136.2 C 219kPa 6:11  
136.3 C 223kPa 6:41  
136.4 C 223kPa 7:11  
136.1 C 222kPa 7:41  
136.4 C 222kPa 8:11  
136.4 C 222kPa 8:11  
VENTING 8:11  
AIR DRYING 9:08  
CYCLE COMPLETE 9:18

-----  
Digital Signature #  
C43CE10EC688C78F  
-----