
Midmark M11 - v1.0.0
Total Cycles: 2041

Name: RITTER 1
Sterilizer ID: V2178198
06/21/2022 01:02 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 109.6 F 0.0
2:00 111.1 F 0.2
4:00 136.4 F 2.1
6:00 170.1 F 4.7
8:00 196.7 F 8.9
10:00 215.6 F 10.4
12:00 224.8 F 6.9
14:00 236.6 F 11.4
16:00 232.9 F 7.8
18:00 244.7 F 12.4
20:00 243.8 F 12.1
22:00 256.8 F 18.8
24:00 268.4 F 26.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.2
1:00 272.5 F 28.6
2:00 272.1 F 28.7
3:00 272.1 F 28.4
4:00 272.3 F 28.8

Min 271.5 F 28.0
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.0 F 0.0
5:00 226.7 F 0.0
10:00 221.0 F 0.0
15:00 213.7 F 0.0
20:00 215.0 F 0.0
25:00 219.7 F 0.0
30:00 218.1 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 25:10
STERILIZING: 4:00
VENTING: 2:57
DRYING: 30:00
TOTAL CYCLE: 01:03:05
CYCLE COMPLETE

06/21/2022 02:05 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____