
Midmark M11 - v1.0.5
Total Cycles: 42

Name: RITTER2
Sterilizer ID: V2313228
05/26/2021 06:35 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.7 F 0.0
2:00 71.0 F 0.2
4:00 73.3 F 0.8
6:00 96.4 F 2.8
8:00 131.0 F 5.8
10:00 159.7 F 9.3
12:00 184.3 F 13.2
14:00 204.2 F 4.1
16:00 214.2 F 7.5
18:00 225.5 F 11.3
20:00 228.7 F 7.1
22:00 240.0 F 12.4
24:00 239.6 F 10.7
26:00 251.5 F 16.6
28:00 262.3 F 22.9
30:00 271.9 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 272.3 F 29.4
2:00 272.0 F 29.7
3:00 271.9 F 29.1
4:00 272.6 F 29.8

Min 271.4 F 28.9
Max 272.6 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.7 F 0.0
5:00 186.4 F 0.0
10:00 187.2 F 0.0
15:00 187.0 F 0.0
20:00 189.6 F 0.0
25:00 190.0 F 0.0
30:00 188.9 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 30:31
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:08:37
CYCLE COMPLETE

05/26/2021 07:43 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____