

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000145  
9:35 14/09/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:31  
PRESSURIZING 4:31  
STERILIZING 5:46  
136.0 C 220kPa 5:46  
136.4 C 226kPa 6:16  
136.4 C 224kPa 6:46  
136.4 C 226kPa 7:16  
136.4 C 226kPa 7:46  
136.4 C 226kPa 8:16  
136.3 C 225kPa 8:46  
136.4 C 226kPa 9:16  
136.4 C 226kPa 9:16  
VENTING 9:16  
AIR DRYING 10:17  
CYCLE COMPLETE 10:21  
-----

Digital Signature #  
1D2E646B38F95EDE  
-----