
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004087
10:17 25/07/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:29
PRESSURIZING 2:32
STERILIZING 3:32
136.4 C 224kPa 3:32
136.3 C 224kPa 4:02
136.4 C 225kPa 4:32
136.2 C 224kPa 5:02
136.3 C 223kPa 5:32
136.2 C 223kPa 6:02
136.4 C 222kPa 6:32
136.4 C 222kPa 7:02
136.4 C 222kPa 7:02
VENTING 7:02
AIR DRYING 7:52
CYCLE COMPLETE 8:45

Digital Signature #
2A6EAC530F009C6F
