
Midmark M11 - v1.0.5
Total Cycles: 3131

Name: RITTER2
Sterilizer ID: V2313228
11/16/2023 09:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.2 F 0.0
2:00 73.8 F 0.0
4:00 77.4 F 0.2
6:00 107.7 F 1.0
8:00 165.8 F 3.1
10:00 202.4 F 6.1
12:00 206.0 F 9.3
14:00 218.9 F 12.8
16:00 222.0 F 5.4
18:00 229.1 F 9.0
20:00 237.9 F 12.9
22:00 235.3 F 9.1
24:00 241.0 F 10.9
26:00 244.6 F 12.7
28:00 255.9 F 18.7
30:00 266.1 F 25.0
32:00 271.6 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 272.3 F 29.4
1:00 272.8 F 29.7
2:00 272.4 F 30.0
3:00 271.6 F 29.4
4:00 272.3 F 29.5

Min 271.5 F 29.2
Max 272.9 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.0 F 0.0
5:00 233.1 F 0.0
10:00 234.0 F 0.0
15:00 228.7 F 0.0
20:00 228.6 F 0.0
25:00 218.6 F 0.0
30:00 233.3 F 0.0

Drying Complete

FILLING: 0:49
HEATING: 32:23
STERILIZING: 4:00
VENTING: 3:28
DRYING: 30:00
TOTAL CYCLE: 01:10:40
CYCLE COMPLETE

11/16/2023 10:54 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____