
Midmark M11 - v1.0.0
Total Cycles: 659

Name: MIDMARK
Sterilizer ID: V2178198
02/10/2021 09:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.0 F 0.0
2:00 188.3 F 2.5
4:00 184.3 F 7.1
6:00 206.7 F 12.7
8:00 226.8 F 8.5
10:00 236.2 F 9.1
12:00 244.6 F 12.5
14:00 242.6 F 11.2
16:00 257.1 F 18.7
18:00 269.4 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.4
1:00 272.1 F 28.8
2:00 272.4 F 28.9
3:00 272.5 F 28.9
4:00 272.7 F 28.9

Min 271.7 F 28.3
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.6 F 0.0
5:00 233.4 F 0.0
10:00 234.8 F 0.0
Dry Cycle Aborted

FILLING: 1:08
HEATING: 19:30
STERILIZING: 4:00
VENTING: 2:55
DRYING: 13:42
TOTAL CYCLE: 00:41:15

02/10/2021 10:21 AM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____