
Midmark M11 - v1.0.0
Total Cycles: 5468

Name: RITTER 1
Sterilizer ID: V2178198
04/02/2025 10:17 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 161.4 F 0.0
2:00 161.2 F 2.2
4:00 164.6 F 5.1
6:00 182.1 F 8.4
8:00 203.0 F 12.1
10:00 205.7 F 4.9
12:00 216.7 F 9.2
14:00 229.7 F 12.8
16:00 232.1 F 9.6
18:00 237.6 F 9.4
20:00 243.6 F 12.9
22:00 254.9 F 18.7
24:00 265.2 F 24.8
26:00 271.9 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.1
1:00 271.9 F 29.4
2:00 272.4 F 29.9
3:00 273.0 F 30.0
4:00 273.1 F 29.9

Min 271.5 F 28.9
Max 273.5 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.0 F 0.0
5:00 193.4 F 0.0
10:00 198.1 F 0.0
15:00 192.4 F 0.0
20:00 202.7 F 0.0
25:00 209.2 F 0.0
30:00 213.4 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 26:08
STERILIZING: 4:00
VENTING: 3:59
DRYING: 30:00
TOTAL CYCLE: 01:05:00
CYCLE COMPLETE

04/02/2025 11:21 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____