

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.7uS / 0.4ppm  
CYCLE NUMBER 000529  
10:54 11/06/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:19  
PRESSURIZING 3:23  
STERILIZING 4:35  
136.0 C 223kPa 4:35  
136.4 C 226kPa 5:05  
136.3 C 233kPa 5:35  
136.3 C 231kPa 6:05  
136.4 C 231kPa 6:35  
136.4 C 231kPa 7:05  
136.1 C 230kPa 7:35  
136.2 C 231kPa 8:05  
136.2 C 231kPa 8:05  
VENTING 8:05  
AIR DRYING 9:00  
CYCLE COMPLETE 24:00  
-----

Digital Signature #  
1C1808A91F53318C  
-----