
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000643
9:58 10/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:29
PRESSURIZING 3:33
STERILIZING 4:47
136.0 C 227kPa 4:47
136.4 C 231kPa 5:17
136.3 C 231kPa 5:47
136.4 C 232kPa 6:17
136.4 C 230kPa 6:47
136.2 C 230kPa 7:17
136.1 C 229kPa 7:47
136.4 C 232kPa 8:17
136.4 C 232kPa 8:17
VENTING 8:18
AIR DRYING 9:11
CYCLE COMPLETE 24:11

Digital Signature #
46ED690AFF251875
