
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001285
12:42 06/10/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:20
PRESSURIZING 4:21
STERILIZING 5:31
136.0 C 220kPa 5:31
136.4 C 224kPa 6:01
136.4 C 223kPa 6:31
136.4 C 223kPa 7:01
136.3 C 224kPa 7:31
136.3 C 224kPa 8:01
136.3 C 223kPa 8:31
136.2 C 223kPa 9:01
136.2 C 223kPa 9:01
VENTING 9:02
AIR DRYING 9:58
CYCLE COMPLETE 24:58

Digital Signature #
06250A750E118453
