
Midmark M11 - v1.0.0
Total Cycles: 5454

Name: RITTER 1
Sterilizer ID: V2178198
03/27/2025 10:47 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 111.6 F 0.0
2:00 113.5 F 0.9
4:00 127.7 F 3.1
6:00 153.9 F 5.8
8:00 175.1 F 9.1
10:00 194.8 F 12.9
12:00 212.2 F 4.9
14:00 215.9 F 9.2
16:00 229.4 F 12.6
18:00 234.5 F 9.4
20:00 240.9 F 10.6
22:00 244.1 F 12.3
24:00 254.5 F 18.1
26:00 265.4 F 24.2
28:00 271.4 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.6
1:00 271.8 F 29.2
2:00 272.7 F 29.6
3:00 271.7 F 29.1
4:00 272.4 F 29.7

Min 270.8 F 28.4
Max 273.5 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.0 F 0.0
5:00 204.8 F 0.0
10:00 206.9 F 0.0
15:00 207.3 F 0.0
20:00 209.5 F 0.0
25:00 213.8 F 0.0
30:00 218.3 F 0.0
Drying Complete

FILLING: 0:44
HEATING: 28:01
STERILIZING: 4:00
VENTING: 3:56
DRYING: 30:00
TOTAL CYCLE: 01:06:41
CYCLE COMPLETE

03/27/2025 11:53 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____