
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002010
8:58 13/07/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:42
PRESSURIZING 3:43
STERILIZING 4:41
136.4 C 220kPa 4:41
136.3 C 223kPa 5:11
136.4 C 219kPa 5:41
136.2 C 223kPa 6:11
136.4 C 222kPa 6:41
136.3 C 224kPa 7:11
136.4 C 223kPa 7:41
136.4 C 224kPa 8:11
136.4 C 224kPa 8:11
VENTING 8:11
AIR DRYING 9:07
CYCLE COMPLETE 9:10

Digital Signature #
4E2B4044050277E9
