
Midmark M11 - v1.0.0
Total Cycles: 2135

Name: RITTER 1
Sterilizer ID: V2178198
07/15/2022 09:22 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 210.2 F 0.0
2:00 203.0 F 3.1
4:00 218.2 F 8.2
6:00 231.4 F 10.4
8:00 234.6 F 8.8
10:00 241.6 F 11.1
12:00 241.7 F 11.4
14:00 239.0 F 9.7
16:00 252.8 F 16.5
18:00 264.1 F 23.2
20:00 271.5 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.1
1:00 272.3 F 28.9
2:00 271.6 F 28.1
3:00 271.5 F 28.6
4:00 272.3 F 28.8

Min 271.1 F 28.0
Max 272.4 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.2 F 0.0
5:00 233.4 F 0.0
10:00 236.2 F 0.0
15:00 223.7 F 0.0
20:00 221.5 F 0.0
25:00 220.2 F 0.0
30:00 222.6 F 0.0
Drying Complete

FILLING: 2:09
HEATING: 20:05
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:59:22
CYCLE COMPLETE

07/15/2022 10:21 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____