

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4367

Name: RITTER 1  
Sterilizer ID: V2178198  
03/07/2024 04:26 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 149.9 F 0.0  
2:00 144.9 F 2.0  
4:00 180.0 F 4.8  
6:00 190.1 F 8.4  
8:00 207.8 F 12.7  
10:00 223.0 F 5.4  
12:00 229.7 F 9.6  
14:00 234.5 F 10.5  
16:00 240.4 F 10.8  
18:00 235.3 F 8.4  
20:00 249.0 F 14.7  
22:00 260.0 F 20.8  
24:00 269.7 F 27.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.8  
1:00 272.3 F 29.2  
2:00 272.0 F 29.3  
3:00 271.9 F 29.0  
4:00 272.7 F 29.7

Min 271.0 F 28.7  
Max 273.1 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.4 F 0.0  
5:00 224.6 F 0.0  
10:00 234.8 F 0.0  
15:00 203.9 F 0.0  
20:00 190.3 F 0.0  
25:00 193.5 F 0.0  
30:00 197.8 F 0.0  
Drying Complete

FILLING: 1:08  
HEATING: 24:58  
STERILIZING: 4:00  
VENTING: 3:47  
DRYING: 30:00  
TOTAL CYCLE: 01:03:53  
CYCLE COMPLETE

03/07/2024 05:29 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_