
Midmark M11 - v1.0.5
Total Cycles: 3848

Name: RITTER2
Sterilizer ID: V2313228
08/12/2024 01:27 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 147.7 F 0.0
2:00 150.5 F 2.3
4:00 166.9 F 5.8
6:00 198.0 F 10.4
8:00 218.8 F 5.4
10:00 233.0 F 9.7
12:00 232.2 F 9.1
14:00 240.8 F 12.4
16:00 240.2 F 11.2
18:00 252.9 F 17.7
20:00 264.5 F 24.5
22:00 271.5 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.2
1:00 271.8 F 29.7
2:00 271.7 F 29.5
3:00 272.4 F 30.0
4:00 271.9 F 29.9

Min 270.9 F 29.1
Max 273.4 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.5 F 0.0
5:00 235.6 F 0.0
10:00 191.6 F 0.0
15:00 186.7 F 0.0
20:00 183.3 F 0.0
25:00 180.7 F 0.0
30:00 178.0 F 0.0
Drying Complete

FILLING: 1:11
HEATING: 22:08
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:00:51
CYCLE COMPLETE

08/12/2024 02:27 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____