
Midmark M11 - v1.0.5
Total Cycles: 3334

Name: RITTER2
Sterilizer ID: V2313228
01/29/2024 04:41 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 170.9 F 0.0
2:00 172.1 F 1.3
4:00 182.3 F 6.6
6:00 208.2 F 13.1
8:00 227.7 F 9.3
10:00 233.1 F 8.0
12:00 238.5 F 9.5
14:00 251.7 F 16.5
16:00 266.1 F 25.0
18:00 265.0 F 24.2

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.6
1:00 272.5 F 28.8
2:00 273.1 F 29.4
3:00 271.8 F 28.7
4:00 272.8 F 29.0

Min 271.8 F 28.4
Max 273.1 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 217.4 F 0.0
Dry Cycle Aborted

FILLING: 1:49
HEATING: 19:26
STERILIZING: 4:00
VENTING: 2:55
DRYING: 0:57
TOTAL CYCLE: 00:29:06

01/29/2024 05:10 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____