

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1248

Name: RITTER2  
Sterilizer ID: V2313228  
07/04/2022 10:47 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 147.4 F 0.0  
2:00 144.7 F 2.2  
4:00 166.1 F 5.6  
6:00 188.1 F 10.1  
8:00 219.3 F 5.3  
10:00 229.7 F 8.6  
12:00 240.5 F 13.5  
14:00 238.7 F 10.6  
16:00 235.1 F 8.6  
18:00 249.4 F 15.2  
20:00 260.9 F 21.7  
22:00 271.4 F 28.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.2  
1:00 272.8 F 29.7  
2:00 271.4 F 29.1  
3:00 272.4 F 29.7  
4:00 271.6 F 29.1

Min 271.3 F 29.0  
Max 273.0 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 189.5 F 0.0  
5:00 232.3 F 0.0  
10:00 219.0 F 0.0  
15:00 211.5 F 0.0  
20:00 195.7 F 0.0  
25:00 200.8 F 0.0  
30:00 196.7 F 0.0  
Drying Complete

FILLING: 1:04  
HEATING: 23:28  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 01:01:40  
CYCLE COMPLETE

07/04/2022 11:48 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_