
Midmark M11 - v1.0.0
Total Cycles: 4600

Name: RITTER 1
Sterilizer ID: V2178198
06/03/2024 07:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.7 F 0.0

C103: HEATUP MODE
STOP PRESSED

** ITEMS NOT STERILE **

06/03/2024 07:45 AM