
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002080
10:21 23/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:00
PRESSURIZING 3:06
STERILIZING 4:21
136.5 C 225kPa 4:21
136.1 C 223kPa 4:51
136.2 C 224kPa 5:21
136.4 C 226kPa 5:51
136.4 C 226kPa 6:21
136.4 C 225kPa 6:51
136.3 C 226kPa 7:21
136.2 C 225kPa 7:51
136.2 C 225kPa 7:51
VENTING 7:51
AIR DRYING 8:48
CYCLE COMPLETE 9:03

Digital Signature #
CFABDF5C6AB080B4
