
Midmark M11 - v1.0.5
Total Cycles: 624

Name: RITTER2
Sterilizer ID: V2313228
12/09/2021 04:19 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 213.9 F 0.0
2:00 206.1 F 4.4
4:00 213.2 F 10.9
6:00 226.3 F 7.1
8:00 240.2 F 13.1
10:00 241.6 F 11.8
12:00 242.1 F 11.5
14:00 256.0 F 18.8
16:00 268.2 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.7
1:00 272.7 F 29.2
2:00 271.6 F 28.8
3:00 272.7 F 29.3
4:00 271.8 F 28.6

Min 271.5 F 28.5
Max 272.9 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.8 F 0.0
5:00 225.1 F 0.0
10:00 230.9 F 0.0
15:00 227.1 F 0.0
20:00 225.5 F 0.0
25:00 229.6 F 0.0
Dry Cycle Aborted

FILLING: 3:41
HEATING: 17:12
STERILIZING: 4:00
VENTING: 3:01
DRYING: 27:31
TOTAL CYCLE: 00:55:25

12/09/2021 05:14 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____