
Midmark M11 - v1.0.0
Total Cycles: 4324

Name: RITTER 1
Sterilizer ID: V2178198
02/22/2024 04:22 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 156.2 F 0.0
2:00 157.8 F 2.8
4:00 169.8 F 6.6
6:00 198.7 F 11.2
8:00 214.3 F 4.5
10:00 226.5 F 9.3
12:00 234.6 F 9.9
14:00 239.5 F 11.3
16:00 237.5 F 9.4
18:00 250.9 F 15.9
20:00 262.3 F 22.4
22:00 272.7 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.5
1:00 272.6 F 29.2
2:00 271.5 F 28.4
3:00 272.4 F 29.3
4:00 272.3 F 28.9

Min 271.3 F 28.3
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.0 F 0.0
5:00 215.7 F 0.0
10:00 218.9 F 0.0
15:00 219.3 F 0.0
20:00 220.1 F 0.0
25:00 221.9 F 0.0
30:00 227.6 F 0.0
Drying Complete

FILLING: 1:16
HEATING: 22:24
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:01:12
CYCLE COMPLETE

02/22/2024 05:23 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____