

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 003328  
14:28 29/03/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:41  
PRESSURIZING 2:43  
STERILIZING 3:50  
136.0 C 220kPa 3:50  
136.4 C 225kPa 4:21  
136.2 C 224kPa 4:51  
136.4 C 224kPa 5:21  
136.2 C 224kPa 5:51  
136.4 C 225kPa 6:21  
136.4 C 225kPa 6:51  
136.3 C 222kPa 7:21  
136.3 C 222kPa 7:21  
VENTING 7:21  
AIR DRYING 8:15  
CYCLE COMPLETE 9:21

-----  
Digital Signature #  
FA0246FC0B2842CC  
-----