
Midmark M11 - v1.0.0
Total Cycles: 3074

Name: RITTER 1
Sterilizer ID: V2178198
04/04/2023 01:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.6 F 0.0
2:00 181.1 F 3.5
4:00 196.8 F 7.7
6:00 213.0 F 12.8
8:00 224.6 F 7.9
10:00 239.3 F 13.3
12:00 239.3 F 10.8
14:00 237.6 F 9.2
16:00 251.8 F 16.2
18:00 264.1 F 23.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 272.5 F 29.1
2:00 271.8 F 28.4
3:00 271.5 F 28.6
4:00 272.5 F 29.0

Min 271.0 F 28.1
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.3 F 0.0
5:00 220.5 F 0.0
10:00 229.4 F 0.0
15:00 231.2 F 0.0
20:00 220.5 F 0.0
25:00 220.2 F 0.0
30:00 219.3 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 19:58
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:58:24
CYCLE COMPLETE

04/04/2023 01:58 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____