

-----  
Midmark M11 - v1.0.5  
Total Cycles: 35

Name: RITTER2  
Sterilizer ID: V2313228  
05/20/2021 08:00 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 74.3 F 0.0  
2:00 76.0 F 0.1  
4:00 87.1 F 1.0  
6:00 105.6 F 2.8  
8:00 135.5 F 5.6  
10:00 165.9 F 9.2  
12:00 187.9 F 13.0  
14:00 203.0 F 4.4  
16:00 215.0 F 7.8  
18:00 226.0 F 11.7  
20:00 229.8 F 7.7  
22:00 239.9 F 12.7  
24:00 239.8 F 10.9  
26:00 251.5 F 16.7  
28:00 262.1 F 22.8  
30:00 271.7 F 29.5

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 29.4  
1:00 272.0 F 29.3  
2:00 272.1 F 29.3  
3:00 272.2 F 29.3  
4:00 272.4 F 29.4

Min 272.0 F 29.2  
Max 272.6 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.6 F 0.0  
5:00 209.0 F 0.0  
10:00 207.7 F 0.0  
15:00 203.6 F 0.0  
20:00 195.7 F 0.0  
25:00 195.7 F 0.0  
30:00 195.3 F 0.0  
Drying Complete

FILLING: 1:01  
HEATING: 30:27  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 01:08:35  
CYCLE COMPLETE

05/20/2021 09:08 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_