
Midmark M11 - v1.0.0
Total Cycles: 4012

Name: RITTER 1
Sterilizer ID: V2178198
11/13/2023 12:48 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 151.9 F 0.0
2:00 154.1 F 3.0
4:00 174.9 F 7.1
6:00 197.1 F 11.6
8:00 216.2 F 5.1
10:00 228.5 F 10.1
12:00 232.1 F 8.1
14:00 241.4 F 12.4
16:00 240.7 F 10.8
18:00 254.1 F 17.5
20:00 265.8 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.5 F 29.2
2:00 272.0 F 28.6
3:00 271.3 F 28.4
4:00 272.4 F 29.2

Min 271.2 F 28.2
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.0 F 0.0
5:00 220.5 F 0.0
10:00 222.4 F 0.0
15:00 223.2 F 0.0
20:00 226.1 F 0.0
25:00 229.2 F 0.0
30:00 225.8 F 0.0
Drying Complete

FILLING: 1:11
HEATING: 21:45
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 01:00:27
CYCLE COMPLETE

11/13/2023 01:48 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____