
Midmark M11 - v1.0.0
Total Cycles: 2973

Name: RITTER 1
Sterilizer ID: V2178198
03/06/2023 10:42 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 192.6 F 0.0
2:00 183.0 F 3.2
4:00 190.7 F 7.6
6:00 217.9 F 12.3
8:00 227.8 F 7.3
10:00 239.9 F 12.6
12:00 237.7 F 10.0
14:00 237.0 F 8.2
16:00 250.1 F 15.2
18:00 262.2 F 22.1
20:00 272.7 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.5 F 28.9
2:00 271.3 F 28.2
3:00 272.6 F 29.0
4:00 271.7 F 28.3

Min 271.3 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.7 F 0.0
5:00 235.0 F 0.0
10:00 221.8 F 0.0
15:00 215.9 F 0.0
20:00 211.1 F 0.0
25:00 211.2 F 0.0
30:00 216.1 F 0.0
Drying Complete

FILLING: 1:16
HEATING: 20:18
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:45
CYCLE COMPLETE

03/06/2023 11:40 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____