
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 006540
8:30 01/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:53
PRESSURIZING 1:55
STERILIZING 2:55
136.5 C 223kPa 2:55
136.3 C 223kPa 3:25
136.2 C 218kPa 3:55
136.4 C 224kPa 4:25
136.4 C 223kPa 4:55
136.2 C 223kPa 5:25
136.2 C 221kPa 5:55
136.4 C 222kPa 6:25
136.4 C 222kPa 6:25
VENTING 6:25
AIR DRYING 7:13
CYCLE COMPLETE 7:18

Digital Signature #
79509C8E542726E4
