
Midmark M11 - v1.0.0
Total Cycles: 5174

Name: RITTER 1
Sterilizer ID: V2178198
12/10/2024 12:02 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 98.8 F 0.0
2:00 103.3 F 0.0
4:00 126.7 F 0.0
6:00 156.0 F 0.1
8:00 183.2 F 0.4
10:00 201.2 F 0.7
12:00 214.1 F 3.7
14:00 223.6 F 6.7
16:00 232.8 F 10.5
18:00 235.5 F 8.8
20:00 241.4 F 11.6
22:00 238.1 F 9.5
24:00 239.0 F 9.3
26:00 249.4 F 14.6
28:00 260.3 F 20.7
30:00 270.2 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.0
1:00 272.4 F 28.6
2:00 272.7 F 28.6
3:00 272.1 F 28.2
4:00 271.6 F 28.1

Min 271.6 F 27.9
Max 272.8 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.5 F 0.0
5:00 187.5 F 0.0
10:00 194.2 F 0.0
15:00 196.2 F 0.0
20:00 196.0 F 0.0
25:00 202.3 F 0.0
30:00 209.9 F 0.0
Drying Complete

FILLING: 0:42
HEATING: 31:02
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 01:09:26
CYCLE COMPLETE

12/10/2024 01:11 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____