

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 007475  
9:53 03/06/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:31  
PRESSURIZING 2:36  
STERILIZING 3:36  
136.0 C 220kPa 3:36  
136.5 C 224kPa 4:06  
136.0 C 222kPa 4:36  
136.4 C 222kPa 5:06  
136.2 C 223kPa 5:36  
136.4 C 223kPa 6:06  
136.3 C 220kPa 6:36  
136.3 C 224kPa 7:06  
136.3 C 224kPa 7:06  
VENTING 7:06  
AIR DRYING 7:57  
CYCLE COMPLETE 8:03

-----  
Digital Signature #  
2577C6039099ECAC  
-----