
Midmark M11 - v1.0.0
Total Cycles: 1110

Name: RITTER 1
Sterilizer ID: V2178198
08/23/2021 11:17 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 143.4 F 0.0
2:00 146.9 F 1.8
4:00 166.9 F 5.3
6:00 197.5 F 10.3
8:00 213.2 F 4.6
10:00 233.1 F 9.9
12:00 232.6 F 8.2
14:00 241.4 F 12.1
16:00 239.3 F 10.2
18:00 252.7 F 16.8
20:00 264.1 F 23.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 271.9 F 28.4
2:00 272.6 F 28.8
3:00 272.6 F 29.1
4:00 271.5 F 28.5

Min 271.5 F 28.3
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 215.0 F 0.0
5:00 228.2 F 0.0
10:00 227.3 F 0.0
15:00 194.6 F 0.0
20:00 182.9 F 0.0
25:00 179.9 F 0.0
30:00 176.9 F 0.0
Drying Complete

FILLING: 1:30
HEATING: 21:54
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:00:31
CYCLE COMPLETE

08/23/2021 12:52 PM
ACCEPTED: X
REJECTED:
Operator: U02