
Midmark M11 - v1.0.5
Total Cycles: 3520

Name: RITTER2
Sterilizer ID: V2313228
04/12/2024 08:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.1 F 0.0
2:00 69.2 F 0.0
4:00 76.3 F 0.0
6:00 98.3 F 0.1
8:00 142.2 F 1.1
10:00 170.4 F 3.8
12:00 197.8 F 7.0
14:00 204.6 F 10.9
16:00 215.2 F 5.9
18:00 223.1 F 8.1
20:00 234.8 F 12.3
22:00 233.3 F 9.2
24:00 239.3 F 10.8
26:00 244.5 F 13.3
28:00 256.0 F 19.4
30:00 266.4 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.3
1:00 272.8 F 30.1
2:00 271.6 F 29.2
3:00 272.4 F 30.2
4:00 272.3 F 29.6

Min 271.3 F 29.2
Max 272.9 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.4 F 0.0
5:00 207.9 F 0.0
10:00 188.4 F 0.0
15:00 184.3 F 0.0
20:00 181.2 F 0.0
25:00 180.9 F 0.0
30:00 181.4 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 31:42
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:09:52
CYCLE COMPLETE

04/12/2024 10:06 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____