
Midmark M11 - v1.0.0
Total Cycles: 2858

Name: RITTER 1
Sterilizer ID: V2178198
02/06/2023 10:06 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.6 F 0.0
2:00 177.1 F 2.9
4:00 193.1 F 6.5
6:00 197.6 F 10.7
8:00 215.6 F 4.3
10:00 226.7 F 9.2
12:00 238.1 F 10.5
14:00 239.5 F 11.1
16:00 236.8 F 8.9
18:00 250.0 F 15.4
20:00 261.5 F 21.9
22:00 272.1 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.7 F 29.3
2:00 271.6 F 28.5
3:00 272.2 F 29.2
4:00 272.4 F 29.1

Min 271.4 F 28.4
Max 272.8 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.3 F 0.0
5:00 202.9 F 0.0
10:00 219.7 F 0.0
15:00 225.9 F 0.0
20:00 227.8 F 0.0
25:00 227.6 F 0.0
30:00 228.9 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 22:33
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:00:46
CYCLE COMPLETE

02/06/2023 11:06 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____