
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000968
10:05 09/03/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:36
PRESSURIZING 2:42
STERILIZING 3:51
136.5 C 233kPa 3:51
136.2 C 230kPa 4:21
136.3 C 232kPa 4:51
136.3 C 231kPa 5:21
136.2 C 231kPa 5:51
136.4 C 232kPa 6:21
136.3 C 232kPa 6:51
136.3 C 232kPa 7:21
136.3 C 232kPa 7:21
VENTING 7:21
AIR DRYING 8:12
CYCLE COMPLETE 8:16

Digital Signature #
115F1A3E5BD99BBA
