
Midmark M11 - v1.0.5
Total Cycles: 3994

Name: RITTER2
Sterilizer ID: V2313228
10/03/2024 11:47 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 146.2 F 0.0
2:00 148.2 F 2.2
4:00 164.0 F 5.4
6:00 187.0 F 9.0
8:00 199.9 F 12.9
10:00 214.0 F 5.8
12:00 225.7 F 9.8
14:00 233.0 F 10.4
16:00 237.3 F 11.0
18:00 235.8 F 9.0
20:00 248.7 F 15.1
22:00 259.1 F 21.1
24:00 269.0 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 29.2
1:00 271.8 F 29.5
2:00 272.9 F 30.0
3:00 272.4 F 29.6
4:00 271.7 F 29.2

Min 271.5 F 29.1
Max 273.0 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.0 F 0.0
5:00 197.1 F 0.0
10:00 199.9 F 0.0
15:00 200.2 F 0.0
20:00 198.1 F 0.0
25:00 196.7 F 0.0
30:00 197.0 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 25:06
STERILIZING: 4:00
VENTING: 3:50
DRYING: 30:00
TOTAL CYCLE: 01:03:60
CYCLE COMPLETE

10/03/2024 12:50 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____